



# Certificate of Compliance

**Certificate:** 80194168 **Master Contract:** 200965  
**Project:** 80194168 **Date Issued:** 2025-11-21  
**Issued to:** **Garrett Electronics Inc.** **Issued by:** *Mengyao Yang*  
1881 W State St *Mengyao Yang*  
Garland, Texas 75042-6797  
United States  
**Attention:** DNU Troy Martin

*The products listed below are eligible to bear the CSA Mark shown with adjacent indicators 'C' and 'US' for Canada and US or with adjacent indicator 'US' for US only or without either indicator for Canada only.*



## **PRODUCTS**

Class 8721 06 ELECTRICAL LABORATORY EQUIPMENT - Electrical Laboratory Equipment  
Class 8721 86 ELECTRICAL LABORATORY EQUIPMENT - Certified to US Standards

### **Electrical test and measurement equipment**

Model(s)	Input Voltage (VAC)	Input Frequency (Hz)	Input Power (W)
Paragon	100-240	50/60	60

### **Notes:**

1. The above model is detachable cord, Equipment Class I, Pollution Degree 2, Overvoltage Category II
2. Mode of operation: Continuous
3. Environmental Conditions: Extended conditions as specified by manufacturer -20°C to 55°C, 95% rH non-condensing max,



**Certificate:** 80194168

**Master Contract:** 200965

**Project:** 80194168

**Date Issued:** 2025-11-21

---

2000m

Conditions of Acceptability:

1. Equipment is only to be installed by manufacturer trained personnel.
2. Equipment is not to be used with flammable liquids.
3. Mains cord-set is not part of this evaluation. The equipment was accepted on the basis that the installation instructions continue to state the use of only an appropriately rated and approved cord-set in accordance with the regulations of the country it is used in.

**APPLICABLE REQUIREMENTS**

CAN/CSA C22.2 No. 61010-1-12, UPD1:2015, UPD2:2016, AMD1:2018, UPD3:2023, UPD4:2024 - Safety requirements for electrical equipment for measurement, control, and laboratory use — Part 1: General requirements - Third Edition

UL 61010-1 3rd ed (Rev. Nov 15, 2024) - UL Standard for Safety Safety Requirements for Electrical Equipment For Measurement, Control, and Laboratory Use; Part 1: General Requirements



**Certificate:** 80194168


**Master Contract:** 200965

**Project:** 80194168

**Date Issued:** 2025-11-21

Notes:

Products certified under Class(es) C872106 have been certified under CSA's ISO/IEC 17065 accreditation with the Standards Council of Canada (SCC). [www.scc.ca](http://www.scc.ca)





## *Supplement to Certificate of Compliance*

**Certificate:** 80194168

**Master Contract:** 200965

*The products listed, including the latest revision described below,  
are eligible to be marked in accordance with the referenced Certificate.*

### **Product Certification History**

---

<b>Project</b>	<b>Date</b>	<b>Description</b>
80194168	2025-11-21	cCSAus certification of Walk Through Metal Detector, model Paragon - based on CB report GARR0105-R0 (CB Cert. # GB-EMT 2341)